

Sample Photo



Test Report

Number: GZHT02276507-S1

Applicant:

Date: Apr 16, 2020
THIS IS TO SUPERSEDE REPORT
NO. GZHT02276507 DATED APR
14, 2020


Sample Description:

One (1) Group Of Submitted Sample Said To Be Makmask Fabric Mask In White With Black Dot.

Applicant's Provided Care Instruction/Label: -

Date Received : Apr 01, 2020
Date Tests Started : Apr 01, 2020
Order No. : -
Style No. : -
Season : -
Buyer's Name : -
Brand Name : Makmask
Vendor's Name : -
Manufacturer's Name : -
Document Accordance : GB 2626-2006 Respiratory Protective Equipment Non-Powered Air-Purifying Particle Respirator
Remark : Sterilization Or Not: Not

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch



Lin Lin
General Manager



Page 1 Of 3

BU / ziminlee

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China
中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房
101、
E201、E301、E401、E501、E601、E701、E801
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

3/F, Hengyun Building, 235 Kaifa Ave., Guangzhou
Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区开发大道235号恒运大厦3楼
Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report


Number: GZHT02276507-S1

Conclusion:

Exhalation Resistance M
Inhalation Resistance M

Note: M = Meet Standard's Requirement F = Below Standard's Requirement
C = Conform To The Declared Fibre Content * = See Remark
= No Comment N/A = Not Applicable

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch



Lin Lin
General Manager



BU / ziminlee

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China
中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房
101、
E201、E301、E401、E501、E601、E701、E801
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663



3/F, Hengyun Building, 235 Kaifa Ave., Guangzhou
Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区开发大道235号恒运大厦3楼
Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

Test Report

Number: GZHT02276507-S1

Tests Conducted (As Requested By The Applicant)

1 Exhalation Resistance (GB 2626 – 2006, Clause 6.6, Head Model Size: M)

		Requirement ≤ 250Pa
Exhalation Resistance (Pa)	As Received	After Temperature and Humidity Pretreatment
Specimen 1	48.8	66.4
Specimen 2	47.5	68.5

2 Inhalation Resistance (GB 2626 – 2006, Clause 6.5, Head Model Size: M)

		Requirement ≤ 350Pa
Inhalation Resistance (Pa)	As Received	After Temperature and Humidity Pretreatment
Specimen 1	49.9	71.7
Specimen 2	48.1	73.9

End of Report

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.

BU / ziminlee

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, 3/F, Hengyun Building, 235 Kaifa Ave., Guangzhou
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China Economic & Technological Development District, Guangzhou, China
中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房 中国广州经济技术开发区开发大道235号恒运大厦3楼
101、
E201、E301、E401、E501、E601、E701、E801 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730
Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663



intertek

Total Quality. Assured.

To :
Attention :

Date : Apr 16, 2020


Re : Report Revision Notification

Labtest Report Number GZHT02276507 date APR 14, 2020

Please be informed that all the content recorded in the above captioned report will be void. This captioned report is now superseded by a revised Labtest Report, Number GZHT02276507-S1 , issued on Apr 16, 2020 .

Thank you for your attention

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch



Lin Lin
General Manager

Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch

3/F, Hengyun Building, 235 Kaifa Ave., Guangzhou
Economic & Technological Development District, Guangzhou, China

深圳天祥质量技术服务有限公司广州开发区分公司

中国广州经济技术开发区开发大道 235 号恒运大厦 3 楼

Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

www.intertek.com